

In Reference To Case No(s): 45671 SDG: MF6R05 (I-0679)

**Contract Laboratory Program
REGIONAL/LABORATORY COMMUNICATION SYSTEM**

Resubmission Request

Laboratory Name:	BON
Lab Contact:	Chris Bonner
Region:	6
Regional Contact:	Raymond Flores - EPA
ESAT Reviewer:	Sonya Meekins - ESAT

In reference to data for the following fractions:

ICP-MS

Summary of Questions/Issues:

The non-blank standard at or below the MA 2542.1 CRQL for thallium was not reported on Form 15 as required by Section II of MA 2542.1. Please make the necessary corrections and resubmissions with the pages properly paginated.

NOTE: Any submitted laboratory resubmission should be clearly marked as "Additional Data" with a cover letter included describing what data is being delivered, which Case the data pertains, and who requested the data (ISM02.3, p. B-9, sec. 2.2.1). Custody seals are required only for regular mail shipments.

Please respond to the above item **within 5 business days (ISM02.3, p. B-9, sec. 2.2.1)** by e-mail to Flores.Raymond@epa.gov. If you have any questions, please contact Mr. Flores at 281-983-2139.

Distribution: (1) Lab Copy, (2) Region Copy, and (3) ESAT Copy